

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 8/21/2014

DBH
mmB

Instrument Serial Number: 010461	As of 21-Aug-14
---	------------------------

Start Date	End Date	Agency Name
20-Aug-14		DFS Central Lab
18-Jun-14	20-Aug-14	Franklin County SO
29-May-14	18-Jun-14	DFS Central Lab
26-Jul-12	29-May-14	Appomattox County SO
10-Jul-12	26-Jul-12	DFS Central Lab
09-Dec-11	10-Jul-12	Henrico County PD
09-Dec-11	09-Dec-11	DFS Central Lab
17-Aug-10	09-Dec-11	Henrico County PD
25-Mar-09	17-Aug-10	DFS Central Lab
28-Oct-08	25-Mar-09	Russell County SO
08-May-08	28-Oct-08	DFS Central Lab

DBH
mmB

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number **010461** Worksheet Start Date **8/20/2014**
Location **Franklin CO SO**
Address **40 E Court St Rocky Mount VA 24151**
DFS Technician **Donald Hall** License No. **21250**

☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☐

Instrument Barometer (mm HG) **729** Reference Barometer (mm HG) **729**
Reference Barometer(RB)Serial # **009109** RB Calibration Due **7/28/2015**

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.290
0.300	0.288	0.279	0.296	Sample 2	0.290
Precision		sample min	sample max	Sample 3	0.289
0.001		0.289	0.290		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.096
0.100	0.096	0.093	0.099	Sample 2	0.096
Precision		sample min	sample max	Sample 3	0.096
0		0.096	0.096		


Dry gas standard Lot No. (with tank no.) **AG332301-27**

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☒ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Removed due to communication issues.

Instrument Serial Number

010461

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Donald B. Hall

Date

8/21/14

Issuing Analyst

Michelle Buller

Date

*8/25/14**TLN*

DBH
mmB




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 08/20/2014
TEST EQUIPMENT NUMBER 010461		

RESULTS: TIME SAMPLE TAKEN 17:26 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 21 Aug 2014 at 09:46

DBH
mmB

Test Results

Instrument Serial Number 010461

Test # 002406

Subject Test

Test Location 1 Department of
Test Date 20 Aug 2014

Test Location 2 Forensic Science
Test Time 17:20
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab
Card Serial Number 121250
Subject's Last Name INSTRUMENT
Subject's Middle Initial
Driver's License Number

Operator's First Name DONALD
Operator's Middle Initial B
License Number 21250
Effective Date 10/01/2012
Expiration Date 10/01/2014
Subject's First Name TEST
Subject's Date of Birth 00/00/0000
Subject's Sex Male

Driver's License Expiration 00/00/0000

Court Name DFS

End Date 20 Aug 2014 End Time 17:28

Result Time 17:26

Result Date 20 Aug 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 17:20

Data Type BLK

Sample Value 0.000

Sample Time 17:21

Data Type CHK

Sample Value 0.097

Sample Time 17:22

Data Type BLK

Sample Value 0.000

Sample Time 17:23

Data Type SUBJ

Sample Value 0.000

Sample Time 17:24

Data Type BLK

Sample Value 0.000

Sample Time 17:25

Data Type SUBJ

Sample Value 0.000

Sample Time 17:26

Data Type BLK

Sample Value 0.000

Sample Time 17:27

Standard Type Dry Gas Std

Standard Value 0.096

Standard Lot Number AG332301-27

Standard Expiration Date 11/19/2015

Tank Pressure 561

Barometric Pressure 729 mmHg

Blow Sample Number 1 Blow Duration 3.89 sec

Blow Volume 2066 cc End-of-Blow Time 17:24

Blow Sample Number 2 Blow Duration 3.92 sec

Blow Volume 1982 cc End-of-Blow Time 17:26

Tamper Evident Stamp 7a106e74

Test Status Code 0

Test Status Success

INTOX EC/IR-11: Accuracy Check

DBH
MMB

Department of Forensic Science

Serial Number: 010461 Test Number: 2407

Test Date: 08/20/2014 Test Time: 17:29 EDT

Dry Gas Target: 0.096

Lot Number: AG332301-21 Exp Date: 11/19/2015

Tank Pressure: 561 psi Barometric Pressure: 729 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	17:29
CHK	0.096	17:30
BLK	0.000	17:32
CHK	0.096	17:32
BLK	0.000	17:34
CHK	0.096	17:34

Test Status: Success

Calibration CRC: 69B6385F

Intox EC/IR-11: Accuracy Check

DBI
mmB

Department of Forensic Science

Serial Number: 010461 Test Number: 2408

Test Date: 08/20/2014 Test Time: 17:41 EDT

Dry Gas Target: 0.288

Lot Number: AG421201-06 Exp Date: 07/31/2016

Tank Pressure: 389 psi Barometric Pressure: 729 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	17:42
CHK	0.290	17:42
BLK	0.000	17:44
CHK	0.290	17:45
BLK	0.000	17:46
CHK	0.289	17:47

Test Status: Success

Calibration CRC: 69B6385F